



People's Primary Healthcare Initiative,
Balochistan

CORRIGENDUM

With reference to the advertisement issued by PPHI-Balochistan for Tender **No. PPHI-B/H.O BLN/11-25/BHUs-Medicines/2025-26/01** dated **20th November, 2025**, the following amendments are hereby notified for information of all prospective bidders:

1. Correction in Specification – IV Infusion Set

Revised Specification in Surgical Item: (Sterile, disposable, non-pyrogenic IV infusion set (with or without air vent), made of medical-grade PVC, 150 cm tube with flow regulator, drip chamber (15–20 drops/mL), latex-free injection port, luer lock connector, DRAP certified, with at least 3 years shelf life). For reference, please check Page No. 8, Serial No. 3 of the tender document

2. Preference for ISO-Certified and FDA-Approved Items

Preference for items complying with ISO 9001, ISO 7886, ISO 8536, and FDA (USA) approval where applicable. Valid certification documents must be attached with the technical bid. For reference, please check page No. 3 of the tender document.

3. General Instructions

All other terms and conditions of the original tender document remain unchanged. This Corrigendum shall form an integral part of the tender document.

**Head Office PPHI-Balochistan H No 8-A Jinnah Town, Samungli Road Quetta Ph No. 081-2863623,
081-2863620, Fax-0812863642**